



**COLLEGE of AMERICAN
PATHOLOGISTS**

325 Waukegan Road, Northfield, Illinois 60093-2750
800-323-4040 • cap.org

ORIGINAL EVALUATION

VITD-B 2022 25-OH Vitamin D

INSTITUTION: Klinikum der Universitat Munchen
Munich GE 80336

ATTENTION: Tim Kuehnle PhD

CAP NUMBER: 8481165-01 **Kit#** 1

KIT INFORMATION:	Kit ID: 35208143	Kit Mailed: 11/21/2022	Original Evaluation: 1/17/2023	Next Mailing Date: 4/24/2023
-------------------------	----------------------------	----------------------------------	--	--

COPIED TO: CAP

LEGEND: Exception Reason Codes appearing in this evaluation:
<NONE>

TIM KUEHNLE

Reviewed By

18 JAN 2023

Date

The College of American Pathologists recommends that the result of this interlaboratory comparison not be used as a sole criterion for judging the performance of any individual clinical laboratory.



**COLLEGE of AMERICAN
PATHOLOGISTS**
325 Waukegan Road, Northfield, Illinois 60093-2750
800-323-4040 • cap.org

CAP Number: 8481165-01 Kit# 1
Institution: Klinikum der Universitat Munchen
Attention: Tim Kuehnle PhD
City / State: Munich GE 80336

Kit ID: 35208143
Kit Mailed: 11/21/2022
Original Evaluation: 1/17/2023

EVALUATION
ORIGINAL

VITD-B 2022 25-OH Vitamin D

Test Unit of Measure Peer Group	Evaluation and Comparative Method Statistics									Plot of the Relative Distance of Your Results from Target as Percentages of allowed Deviation Survey -100-----Mean-----+100
	Specimen	Your Result	Mean	S.D.	No. of Labs	S.D.I	Limits of Acceptability Lower Upper		Your Grade	
25 OH Vitamin D ng/mL IDS i10, iSYS	VITD-04				6					
	VITD-05				6					
	VITD-06				6					
All Methods	VITD-04	65.1	67.64	8.43	2113	-0.3	50.7	84.6	Acceptable	
	VITD-05	37.4	38.20	5.41	2141	-0.1	28.6	47.8	Acceptable	
	VITD-06	18.6	17.70	2.77	2137	+0.3	12.6	22.7	Acceptable	

The College of American Pathologists recommends that the result of this interlaboratory comparison not be used as a sole criterion for judging the performance of any individual clinical laboratory.