



**COLLEGE of AMERICAN  
PATHOLOGISTS**

325 Waukegan Road, Northfield, Illinois 60093-2750  
800-323-4040 • cap.org

**ORIGINAL EVALUATION**

**TM-B 2025 Tumor Markers (TM)**

**INSTITUTION:** Klinikum der Universitat Munchen  
Munich GE 80336

**ATTENTION:** Tim Kuehnle PhD

**CAP NUMBER:** 8481165-01      **Kit Sequence:** 1

<b>KIT INFORMATION:</b>	<b>Kit #:</b>	<b>Kit Mailed:</b>	<b>Original Evaluation:</b>	<b>Next Mailing Date:</b>
	38564247	7/14/2025	8/18/2025	1/5/2026

**COPIED TO:** CAP

**LEGEND:** **Exception Reason Codes appearing in this evaluation:**  
[20] = Response was not formally graded due to insufficient peer group data. Please see the participant summary for additional information.

Reviewed By

Date

The College of American Pathologists recommends that the result of this interlaboratory comparison not be used as a sole criterion for judging the performance of any individual clinical laboratory.



**COLLEGE of AMERICAN  
PATHOLOGISTS**  
325 Waukegan Road, Northfield, Illinois 60093-2750  
800-323-4040 • cap.org

CAP Number: 8481165-01      Kit Sequence 1  
Institution: Klinikum der Universitat Munchen  
Attention: Tim Kuehnle PhD  
City / State: Munich GE 80336

Kit #: 38564247  
Kit Mailed: 7/14/2025  
Original Evaluation: 8/18/2025

**EVALUATION  
ORIGINAL**

**TM-B 2025 Tumor Markers (TM)**

Test Unit of Measure Peer Group	Evaluation and Comparative Method Statistics									Plot of the Relative Distance of Your Results from Target as Percentages of allowed Deviation Survey    -100-----Mean-----+100
	Specimen	Your Result	Mean	S.D.	No. of Labs	S.D.I	Limits of Acceptability Lower    Upper		Your Grade	
ACTH pg/mL DIASORIN	TM-04	48.4	48.04	7.91	22	0.0	24.2	71.8	Acceptable	
	TM-05	27.9	29.94	8.25	22	-0.2	5.1	54.8	Acceptable	
	TM-06	161.0	164.69	27.48	22	-0.1	82.2	247.2	Acceptable	
Thyroglobulin ng/mL OTHER	TM-04	22.2			22				See Note [20]	
	TM-05	0.8			22				See Note [20]	
	TM-06	38.3			22				See Note [20]	

The College of American Pathologists recommends that the result of this interlaboratory comparison not be used as a sole criterion for judging the performance of any individual clinical laboratory.