



**COLLEGE of AMERICAN
PATHOLOGISTS**
325 Waukegan Road, Northfield, Illinois 60093-2750
800-323-4040 • cap.org

ORIGINAL EVALUATION

TM-B 2021 Tumor Markers

INSTITUTION:

Klinikum der Universitat Munchen
Munich GE 80336

ATTENTION:

Tim Kuehne PhD

CAP NUMBER:

8481165-01

Kit# 1

KIT INFORMATION:

Kit ID: 33828201 **Kit Mailed:** 7/19/2021

Original Evaluation: 8/19/2021

Next Mailing Date: 1/3/2022

COPIED TO:

CAP

LEGEND:

Exception Reason Codes appearing in this evaluation:
[20] = Response was not formally graded due to insufficient peer group data. Please see the participant summary for additional information.

Reviewed By  Date 

The College of American Pathologists recommends that the result of this interlaboratory comparison not be used as a sole criterion for judging the performance of any individual clinical laboratory.



**COLLEGE of AMERICAN
PATHOLOGISTS**
325 Mantegan Road, Northfield, Illinois 60063-2750
800-323-4040 • cap.org

CAP Number: 8481165-01 **Kit#:** 1
Institution: Klinikum der Universität München
Attention: Tim Kuehnl PhD
City / State: Munich GE 80336

Kit ID: 33828201
Kit Mailed: 7/19/2021
Original Evaluation: 8/19/2021

**EVALUATION
ORIGINAL**

TM-B 2021 Tumor Markers

Test Unit of Measure Peer Group	Evaluation and Comparative Method Statistics										Plot of the Relative Distance of Your Results from Target as Percentages of allowed Deviation Survey
	Specimen	Your Result	Mean	S.D.	No. of Labs	S.D.I	Lower	Upper	Your Grade		
ACTH pg/mL DIASORIN	TM-04	227.6	194.35	18.05	12	+1.8	140.2	248.5	Acceptable		
	TM-05	17.0	16.10	1.74	11	+0.5	10.8	21.4	Acceptable		
	TM-06	7.1	8.34	2.58	10	-0.5	0.5	16.1	Acceptable		
Thyroglobulin ng/mL OTHER	TM-04	56.9			14				See Note [20]		
	TM-05	29.3			13				See Note [20]		
	TM-06	2.3			12				See Note [20]		

The College of American Pathologists recommends that the result of this interlaboratory comparison not be used as a sole criterion for judging the performance of any individual clinical laboratory.