



**COLLEGE of AMERICAN
PATHOLOGISTS**

325 Waukegan Road, Northfield, Illinois 60093-2750
800-323-4040 • cap.org

ORIGINAL EVALUATION

Y-A 2026 Sex Hormones (Y)

INSTITUTION: Klinikum der Universitat Munchen
Munich GE 80336

ATTENTION: Tim Kuehnle PhD

CAP NUMBER: 8481165-01 **Kit Sequence:** 1

KIT INFORMATION:	Kit #: 40068390	Kit Mailed: 4/20/2026	Original Evaluation: 6/5/2026	Next Mailing Date: 8/10/2026
-------------------------	---------------------------	---------------------------------	---	--

COPIED TO: CAP

LEGEND: **Exception Reason Codes appearing in this evaluation:**
[20] = Response was not formally graded due to insufficient peer group data. Please see the participant summary for additional information.

Reviewed By

Date

The College of American Pathologists recommends that the result of this interlaboratory comparison not be used as a sole criterion for judging the performance of any individual clinical laboratory.



**COLLEGE of AMERICAN
PATHOLOGISTS**

325 Waukegan Road, Northfield, Illinois 60093-2750
800-323-4040 • cap.org

CAP Number: 8481165-01 Kit Sequence 1

Institution: Klinikum der Universität München

Attention: Tim Kuehne PhD

City / State: Munich GE 80336

Kit #: 40068390

Kit Mailed: 4/20/2026

Original Evaluation: 6/5/2026

**EVALUATION
ORIGINAL**

Y-A 2026 Sex Hormones (Y)

Test Unit of Measure Peer Group	Evaluation and Comparative Method Statistics									Plot of the Relative Distance of Your Results from Target as Percentages of allowed Deviation Survey -100-----Mean-----+100
	Specimen	Your Result	Mean	S.D.	No. of Labs	S.D.I	Limits of Acceptability Lower Upper		Your Grade	
Growth Hormone ng/mL (ug/L) IDS i10, iSYS	Y-01	12.7	12.12	0.45	11	+1.3	10.7	13.5	Acceptable	
	Y-02	0.4	0.40	0.04	12	0.0	0.2	0.6	Acceptable	
	Y-03	4.3	4.08	0.16	12	+1.3	3.5	4.6	Acceptable	
	Y-04	4.6	4.49	0.23	12	+0.5	3.8	5.2	Acceptable	
	Y-05	8.5	8.23	0.31	12	+0.8	7.2	9.2	Acceptable	
IGF-1 ng/mL IDS i10, iSYS	Y-01	103.0	101.09	6.53	17	+0.3	81.4	120.7	Acceptable	
	Y-02	10.0			2				See Note [20]	
	Y-03	42.0	39.69	3.31	17	+0.7	29.7	49.7	Acceptable	
	Y-04	42.0	41.64	2.69	17	+0.1	33.5	49.8	Acceptable	
	Y-05	72.0	69.30	4.51	17	+0.6	55.7	82.9	Acceptable	

The College of American Pathologists recommends that the result of this interlaboratory comparison not be used as a sole criterion for judging the performance of any individual clinical laboratory.