



ORIGINAL EVALUATION

TM-B 2020 Tumor Markers

INSTITUTION:

Klinikum der Universitat Munchen
Munich GE 80336
Tim Kuehnle PhD

ATTENTION:

CAP NUMBER:

8481165-01

KIT INFORMATION:

Kit ID:
32786358

Kit Mailed:
7/20/2020

Original Evaluation:
8/26/2020

Next Mailing Date:
1/4/2021

COPIED TO:

CAP

LEGEND:

Exception Reason Codes appearing in this evaluation

[20] = Response was not formally graded due to insufficient peer group data. Please see the participant summary for additional information.


Tim Kuehnle

Reviewed By

07 SEP 2020

Date

The College of American Pathologists recommends that the result of this interlaboratory comparison not be used as a sole criterion for judging the performance of any individual clinical laboratory.



**COLLEGE of AMERICAN
PATHOLOGISTS**
325 Waukegan Road, Northfield, Illinois 60093-2750
800-323-4040 • cap.org

CAP Number: 8481165-01 **Kit#** 1

Institution: Klinikum der Universität München

Attention: Tim Kuehnle PhD

City / State: Munich GE 80336

Kit ID: 32786358

Kit Mailed: 7/20/2020

Original Evaluation: 8/26/2020

EVALUATION ORIGINAL

TM-B 2020 Tumor Markers

Test Unit of Measure Peer Group	Evaluation and Comparative Method Statistics						Plot of the Relative Distance of Your Results from Target as Percentages of allowed Deviation Survey -100-----Mean-----+ 100
	Specimen	Your Result	Mean	S.D.	No. of Labs	Limits of Acceptability S.D.I Lower Upper	Your Grade
ACTH pg/mL DIASORIN	TM-04	36.1	38.16	2.66	11	-0.8 30.1	46.2 Acceptable
	TM-05	25.9	29.74	3.64	11	-1.1 18.8	40.7 Acceptable
	TM-06	205.5	207.09	14.16	11	-0.1 164.6	249.6 Acceptable
Thyroglobulin ng/mL OTHER	TM-04	27.5			14		See Note [20]
	TM-05	2.3			14		See Note [20]
	TM-06	58.4			13		See Note [20]

