



**COLLEGE of AMERICAN  
PATHOLOGISTS**  
325 Waukegan Road, Northfield, Illinois 60093-2750  
800-323-4060 • cap.org

**ORIGINAL EVALUATION**

**T1M-B 2024 Tumor Markers**

**INSTITUTION:** Klinikum der Universität München

Munich GE 80336

**ATTENTION:** Tim Kuehnle PhD

**CAP NUMBER:** 8481165-01

**Kit#** 1

**KIT INFORMATION:**

**Kit ID:** 37625585

**Kit Mailed:** 7/15/2024

**Original Evaluation:** 8/15/2024

**Next Mailing Date:** 1/6/2025

**COPIED TO:**

CAP

**LEGEND:**

**Exception Reason Codes appearing in this evaluation:**

[20] = Response was not formally graded due to insufficient peer group data. Please see the participant summary for additional information.

Reviewed By 

Date  16. Oct 2024

The College of American Pathologists recommends that the result of this interlaboratory comparison not be used as a sole criterion for judging the performance of any individual clinical laboratory.



**COLLEGE of AMERICAN  
PATHOLOGISTS**  
325 Waukegan Road, Northfield, Illinois 60093-2750  
800-323-4040 • cap.org

CAP Number: 8481165-01      Kit# 1  
Institution: Klinikum der Universität München  
Attention: Tim Kuehle PhD  
City / State: Munich GE 80336

Kit ID: 37625585  
Kit Mailed: 7/15/2024  
Original Evaluation: 8/15/2024

**EVALUATION  
ORIGINAL**

**TM-B 2024 Tumor Markers**

Test Unit of Measure Peer Group	Evaluation and Comparative Method Statistics										Plot of the Relative Distance of Your Results from Target as Percentages of Allowed Deviation Survey
	Specimen	Your Result	Mean	S.D.	No. of Labs	S.D.I	Limits of Acceptability Lower	Upper	Your Grade		
ACTH pg/mL DIASORIN	TM-04	12.5	15.67	2.01	16	-1.6	9.6	21.8	Acceptable		
	TM-05	151.7	178.46	16.05	15	-1.7	130.3	226.7	Acceptable		
	TM-06	45.2	55.49	7.08	16	-1.5	34.2	76.8	Acceptable		
Thyroglobulin ng/mL OTHER	TM-04	0.4			16				See Note [20]		
	TM-05	35.6			16				See Note [20]		
	TM-06	15.6			16				See Note [20]		

The College of American Pathologists recommends that the result of this interlaboratory comparison not be used as a sole criterion for judging the performance of any individual clinical laboratory.