



**COLLEGE of AMERICAN  
PATHOLOGISTS**  
326 Waukegan Road, Northfield, Illinois 60093-2750  
800-323-4040 • cap.org

**ORIGINAL EVALUATION**

**LS-A 2024 Ligand Special**

**INSTITUTION:**

Klinikum der Universität München  
München GE 80336

**ATTENTION:**

Tim Kuehnle PhD

**CAP NUMBER:**

8481165-01

**Kit#** 2

**KIT INFORMATION:**

**Kit ID:**  
37622616

**Kit Mailed:**  
5/20/2024

**Original Evaluation:**  
6/19/2024

**Next Mailing Date:**  
9/16/2024

**COPIED TO:**

CAP

**LEGEND:**

Exception Reason Codes appearing in this evaluation:  
<NONE>



Reviewed By Tim Kuehnle

Date 25 Jun 2024

The College of American Pathologists recommends that the result of this interlaboratory comparison not be used as a sole criterion for judging the performance of any individual clinical laboratory.

**EVALUATION  
ORIGINAL**

**LS-A 2024 Ligand Special**

Test Unit of Measure Peer Group	Evaluation and Comparative Method Statistics										Plot of the Relative Distance of Your Results from Target as Percentages of allowed Deviation Survey -100-----Mean-----+100
	Specimen	Your Result	Mean	S.D.	No. of Labs	S.D.I	Limits of Acceptability Lower	Upper	Your Grade		
Aldosterone ng/dL DIASORIN LIAISON	RAP-01	77.90	87.618	8.254	88	-1.2	62.85	112.39	Acceptable		
	RAP-02	38.90	43.014	3.720	90	-1.1	31.85	54.18	Acceptable		
	RAP-03	12.40	13.434	1.552	92	-0.7	8.77	18.09	Acceptable		
Renin, direct DIASORIN LIAISON uU/mL	RAP-01	100.4	95.37	16.35	43	+0.3	46.3	144.5	Acceptable		
	RAP-02	129.5	129.03	27.61	42	0.0	46.1	211.9	Acceptable		
	RAP-03	13.7	12.93	1.16	43	+0.7	9.4	16.5	Acceptable		

The College of American Pathologists recommends that the result of this interlaboratory comparison not be used as a sole criterion for judging the performance of any individual clinical laboratory.