



**COLLEGE of AMERICAN
PATHOLOGISTS**

325 Waukegan Road, Northfield, Illinois 60093-2750
800-323-4040 • cap.org

ORIGINAL EVALUATION

VITD-A 2025 25-OH Vitamin D

INSTITUTION: Klinikum der Universitat Munchen
Munich BY GE 80336

ATTENTION: Tim Kuehnle PhD

CAP NUMBER: 8481165-01 **Kit#** 1

KIT INFORMATION:	Kit ID: 38564090	Kit Mailed: 4/22/2025	Original Evaluation: 5/21/2025	Next Mailing Date: 11/17/2025
-------------------------	----------------------------	---------------------------------	--	---

COPIED TO: CAP

LEGEND: Exception Reason Codes appearing in this evaluation:
<NONE>

Reviewed By

Date

The College of American Pathologists recommends that the result of this interlaboratory comparison not be used as a sole criterion for judging the performance of any individual clinical laboratory.



**COLLEGE of AMERICAN
PATHOLOGISTS**
325 Waukegan Road, Northfield, Illinois 60093-2750
800-323-4040 • cap.org

CAP Number: 8481165-01 **Kit#** 1
Institution: Klinikum der Universitat Munchen
Attention: Tim Kuehnle PhD
City / State: Munich BY GE 80336

Kit ID: 38564090
Kit Mailed: 4/22/2025
Original Evaluation: 5/21/2025

**EVALUATION
ORIGINAL**

VITD-A 2025 25-OH Vitamin D

Test Unit of Measure Peer Group	Evaluation and Comparative Method Statistics									Plot of the Relative Distance of Your Results from Target as Percentages of allowed Deviation Survey -100-----Mean-----+100
	Specimen	Your Result	Mean	S.D.	No. of Labs	S.D.I	Limits of Acceptability Lower Upper		Your Grade	
25 OH Vitamin D ng/mL IDS i10, iSYS	VITD-01				6					
	VITD-02				6					
	VITD-03				6					
<hr/>										
All Methods	VITD-01	61.2	73.65	7.54	2311	-1.7	55.2	92.1	Acceptable	
	VITD-02	40.7	43.51	4.60	2310	-0.6	32.6	54.4	Acceptable	
	VITD-03	14.6	14.36	2.23	2273	+0.1	9.3	19.4	Acceptable	

The College of American Pathologists recommends that the result of this interlaboratory comparison not be used as a sole criterion for judging the performance of any individual clinical laboratory.