



COLLEGE of AMERICAN
PATHOLOGISTS

325 Waukegan Road, Northfield, Illinois 60093-2750
800-323-4040 • cap.org

ORIGINAL EVALUATION

Y-B 2026 Sex Hormones (Y)

INSTITUTION: Klinikum der Universitat Munchen
Munich GE 80336

ATTENTION: Tim Kuehnle PhD

CAP NUMBER: 8481165-01 Kit Sequence: 1

KIT INFORMATION:	Kit #:	Kit Mailed:	Original Evaluation:	Next Mailing Date:
	40068362	8/10/2026	9/18/2026	12/14/2026

COPIED TO: CAP

LEGEND: Exception Reason Codes appearing in this evaluation:
<NONE>

Reviewed By

Date

The College of American Pathologists recommends that the result of this interlaboratory comparison not be used as a sole criterion for judging the performance of any individual clinical laboratory.



**COLLEGE of AMERICAN
PATHOLOGISTS**
325 Waukegan Road, Northfield, Illinois 60093-2750
800-323-4040 • cap.org

CAP Number: 8481165-01 **Kit Sequence 1**
Institution: Klinikum der Universitat Munchen
Attention: Tim Kuehnle PhD
City / State: Munich GE 80336

Kit #: 40068362
Kit Mailed: 8/10/2026
Original Evaluation: 9/18/2026

EVALUATION
ORIGINAL

Y-B 2026 Sex Hormones (Y)

Test Unit of Measure Peer Group	Evaluation and Comparative Method Statistics									Plot of the Relative Distance of Your Results from Target as Percentages of allowed Deviation
	Specimen	Your Result	Mean	S.D.	No. of Labs	S.D.I	Limits of Acceptability		Your Grade	
							Lower	Upper		Survey -100-----Mean-----+100
Growth Hormone ng/mL (ug/L) IDS i10, iSYS	Y-06	9.9	9.44	0.35	11	+1.3	8.3	10.5	Acceptable	
	Y-07	5.6	5.02	0.29	11	+2.0	4.1	5.9	Acceptable	
	Y-08	21.4	19.69	0.90	11	+1.9	16.9	22.4	Acceptable	
	Y-09	15.6	14.30	0.70	11	+1.9	12.2	16.4	Acceptable	
	Y-10	12.7	12.25	0.44	11	+1.0	10.9	13.6	Acceptable	
IGF-1 ng/mL IDS i10, iSYS	Y-06	22.0	20.40	1.85	16	+0.9	14.8	26.0	Acceptable	
	Y-07	14.0	14.23	1.64	14	-0.1	9.3	19.2	Acceptable	
	Y-08	40.0	33.88	4.04	16	+1.5	21.7	46.0	Acceptable	
	Y-09	30.0	26.21	3.42	16	+1.1	15.9	36.5	Acceptable	
	Y-10	30.0	25.83	4.20	16	+1.0	13.2	38.5	Acceptable	

The College of American Pathologists recommends that the result of this interlaboratory comparison not be used as a sole criterion for judging the performance of any individual clinical laboratory.