



**COLLEGE of AMERICAN
PATHOLOGISTS**

325 Waukegan Road, Northfield, Illinois 60093-2750
800-323-4040 • cap.org

ORIGINAL EVALUATION

TM-A 2024 Tumor Markers

INSTITUTION:

Klinikum der Universitat Munchen
Munich GE 80336

ATTENTION:

Tim Kuehnle PhD

CAP NUMBER:

8481165-01 Kit# 1

KIT INFORMATION:

Kit ID: 37625558
Kit Mailed: 1/16/2024
Original Evaluation: 2/23/2024

Next Mailing Date:
7/15/2024

COPIED TO:

CAP

IMPORTANT NOTES:

LEGEND:

Exception Reason Codes appearing in this evaluation:

[20] = Response was not formally graded due to insufficient peer group data. Please see the participant summary for additional information.

Handwritten signature

Reviewed By

07 MAR 2024

Date

The College of American Pathologists recommends that the result of this interlaboratory comparison not be used as a sole criterion for judging the performance of any individual clinical laboratory.

CAP Number: 8481165-01

Kit# 1

Kit ID: 37625558

Institution: Klinikum der Universität München

Kit Mailed: 1/16/2024

Attention: Tim Kuehnle PhD

Original Evaluation: 2/23/2024

City / State: Munich GE 80336

**EVALUATION
ORIGINAL**

TM-A 2024 Tumor Markers

Test Unit of Measure Peer Group	Evaluation and Comparative Method Statistics							Your Result	Limits of Acceptability			Your Grade	Plot of the Relative Distance of Your Results from Target as Percentages of allowed Deviation Survey -100-----Mean-----+100
	Specimen	Mean	S.D.	No. of Labs	S.D.I	Lower	Upper		Upper	Grade			
ACTH pg/mL DIASORIN	TM-01	47.5	6.24	15	-1.2	36.1	73.6	Acceptable					
	TM-02	14.4	4.17	16	+0.6	0.0	24.5	Acceptable					
	TM-03	161.1	31.39	16	-0.5	81.9	270.3	Acceptable					
Thyroglobulin ng/mL OTHER	TM-01	13.0		28				See Note [20]					
	TM-02	1.3		27				See Note [20]					
	TM-03	25.1		27				See Note [20]					