



**COLLEGE of AMERICAN  
PATHOLOGISTS**

325 Waukegan Road, Northfield, Illinois 60093-2750  
800-323-4040 • cap.org

**ORIGINAL EVALUATION**

**VITD-B 2025 25-OH Vitamin D (VITD)**

**INSTITUTION:** Klinikum der Universitat Munchen  
Munich GE 80336

**ATTENTION:** Tim Kuehnle PhD

**CAP NUMBER:** 8481165-01      **Kit Sequence:** 1

<b>KIT INFORMATION:</b>	<b>Kit #:</b>	<b>Kit Mailed:</b>	<b>Original Evaluation:</b>	<b>Next Mailing Date:</b>
	38564101	11/17/2025	12/15/2025	4/20/2026

**COPIED TO:** CAP

**LEGEND:** Exception Reason Codes appearing in this evaluation:  
<NONE>

Reviewed By

Date

The College of American Pathologists recommends that the result of this interlaboratory comparison not be used as a sole criterion for judging the performance of any individual clinical laboratory.



**COLLEGE of AMERICAN  
PATHOLOGISTS**

325 Waukegan Road, Northfield, Illinois 60093-2750  
800-323-4040 • cap.org

CAP Number: 8481165-01      Kit Sequence 1  
Institution: Klinikum der Universitat Munchen  
Attention: Tim Kuehnle PhD  
City / State: Munich GE 80336

Kit #: 38564101  
Kit Mailed: 11/17/2025  
Original Evaluation: 12/15/2025

**EVALUATION  
ORIGINAL**

**VITD-B 2025 25-OH Vitamin D (VITD)**

Test Unit of Measure Peer Group	Evaluation and Comparative Method Statistics									Plot of the Relative Distance of Your Results from Target as Percentages of allowed Deviation Survey -100-----Mean-----+100
	Specimen	Your Result	Mean	S.D.	No. of Labs	S.D.I	Limits of Acceptability Lower Upper		Your Grade	
25 OH Vitamin D ng/mL IDS i10, iSYS	VITD-04				7					
	VITD-05				7					
	VITD-06				7					
All Methods	VITD-04	38.5	42.60	4.78	2351	-0.9	31.9	53.3	Acceptable	
	VITD-05	47.3	62.60	6.26	2346	-2.4	46.9	78.3	Acceptable	
	VITD-06	17.1	18.24	2.50	2336	-0.5	13.2	23.3	Acceptable	

The College of American Pathologists recommends that the result of this interlaboratory comparison not be used as a sole criterion for judging the performance of any individual clinical laboratory.