



**COLLEGE of AMERICAN
PATHOLOGISTS**
325 Waukegan Road, Northfield, Illinois 60093-2750
800-323-4040 • cap.org

ORIGINAL EVALUATION

TM-A 2022 Tumor Markers

INSTITUTION:

Klinikum der Universitat Munchen
Munich GE 80336

ATTENTION:

Tim Kuehnle PhD

CAP NUMBER:

8481165-01

Kit# 1

KIT INFORMATION:

Kit ID: 35210283
Kit Mated: 1/3/2022

Original Evaluation:

2/24/2022

Next Mating Date:

7/18/2022

COPIED TO:

CAP

LEGEND:

Exception Reason Codes appearing in this evaluation:
[20] = Response was not formally graded due to insufficient peer group data. Please see the participant summary for additional information.

Reviewed By  Date 

The College of American Pathologists recommends that the result of this interlaboratory comparison not be used as a sole criterion for judging the performance of any individual clinical laboratory.



**COLLEGE of AMERICAN
PATHOLOGISTS**
325 Westlaken Road, Northfield, Illinois 60093-2750
800-523-4040 • cap.org

CAP Number: 8481165-01 Kit# 1
Institution: Klinikum der Universität München
Attention: Tim Kuehle PhD
City / State: Munich GE 80336

Kit ID: 35210283
Kit Matld: 1/3/2022
Original Evaluation: 2/24/2022

**EVALUATION
ORIGINAL**

TM-A 2022 Tumor Markers

Test Unit of Measure Peer Group	Specimen	Your Result	Mean	S.D.	No. of Labs	S.D.I	Limits of Acceptability		Your Grade	Plot of the Relative Distance of Your Results from Target as Percentages of allowed Deviation Survey -100-----Mean-----+100
							Lower	Upper		
ACTH pg/mL DIASORIN	TM-01	14.1	10.85	2.42	13	+1.3	3.5	18.1	Acceptable	
	TM-02	18.1	16.04	2.93	13	+0.7	7.2	24.9	Acceptable	
	TM-03	217.5	210.73	31.37	13	+0.2	116.6	304.9	Acceptable	
Thyroglobulin ng/mL OTHER	TM-01	1.2			13				See Note [20]	
	TM-02	26.8			15				See Note [20]	
	TM-03	54.2			16				See Note [20]	

The College of American Pathologists recommends that the result of this interlaboratory comparison not be used as a sole criterion for judging the performance of any individual clinical laboratory.