



LS-A 2025 Ligand Special (RAP)

INSTITUTION: Klinikum der Universitat Munchen
Munich BY GE 80336

ATTENTION: Tim Kuehnle PhD

CAP NUMBER: 8481165-01 Kit Sequence: 2

KIT INFORMATION:	Kit #:	Kit Mailed:	Original Evaluation:	Next Mailing Date:
	38562719	4/22/2025	5/30/2025	9/15/2025

COPIED TO: CAP

LEGEND: Exception Reason Codes appearing in this evaluation:
<NONE>

Reviewed By

Date



**COLLEGE of AMERICAN
PATHOLOGISTS**

325 Waukegan Road, Northfield, Illinois 60093-2750
800-323-4040 • cap.org

CAP Number: 8481165-01 Kit Sequence 2
Institution: Klinikum der Universitat Munchen
Attention: Tim Kuehnle PhD
City / State: Munich BY GE 80336

Kit #: 38562719
Kit Mailed: 4/22/2025
Original Evaluation: 5/30/2025

**EVALUATION
ORIGINAL**

LS-A 2025 Ligand Special (RAP)

Test Unit of Measure Peer Group	Evaluation and Comparative Method Statistics									Plot of the Relative Distance of Your Results from Target as Percentages of allowed Deviation	
	Specimen	Your Result	Mean	S.D.	No. of Labs	S.D.I	Lower	Upper	Your Grade	Survey	-100-----Mean-----+100
Aldosterone ng/dL DIASORIN LIAISON	RAP-01	16.30	15.616	1.661	91	+0.4	10.63	20.60	Acceptable	LS-A 2025 LS-A 2024	
	RAP-02	97.10	92.446	8.387	84	+0.6	67.28	117.61	Acceptable		
	RAP-03	40.40	39.802	3.684	88	+0.2	28.75	50.86	Acceptable		
Renin, direct DIASORIN LIAISON uU/mL	RAP-01	12.2	11.80	1.28	41	+0.3	7.9	15.7	Acceptable	LS-A 2025 LS-A 2024	
	RAP-02	97.5	104.17	16.02	40	-0.4	56.1	152.3	Acceptable		
	RAP-03	125.3	142.25	25.51	40	-0.7	65.7	218.8	Acceptable		

The College of American Pathologists recommends that the result of this interlaboratory comparison not be used as a sole criterion for judging the performance of any individual clinical laboratory.